

D3R Status

Submitted by mvega on Thu, 2013-05-09 23:27

Date:

Thursday, May 9, 2013 - 23:15

Instrument List:

- [Green](#)

Discussion:

D3R operated fairly well today. Scans were stopped briefly to investigate fan failure on one of the thermoelectric coolers on the IF box. After inspection, a small rock was found inside fan cage which may have jammed one of the fans. Spare fans were ordered (shipped to Mike Watson to be delivered on 5/13) just in case the current ones may need to be replaced in the near future.

BB scans performed at: 14:29, 17:51 and 18:11 UTC.

MVega and HChen
